



Local Government
Madhyapur Thimi Municipality


Office of the Municipal Executive,
Madhyapur Thimi, Bhaktapur
Bagmati Province, Nepal

Notice for Bid Amendment no:-4

Date of Bid Publication: 2021/04/20

Bid Amendment no:-4 Publication Date: 2021/05/28

This is to notify all the concerned bidders/suppliers that some points of Amended Technical Specification of 128 slice Computed Tomography (CT) Scanner of IFB no. MTM/ICB/Goods/01/2077/078, Supply, Delivery, Installation, Commissioning and Testing of 128 Slice Computed Tomography (CT) Scan Machine with complete set of accessories at Nepal Korea Friendship Municipality Hospital have been amended. Interested and eligible bidder/suppliers may obtain further information from www.bolpatra.gov.np/egp and www.madhyapurthimimun.gov.np.


कृष्णा गोपाल श्रेष्ठ
इन्जिनियर



Bid Amendment no:-4

IFB No.: MTM/ICB/Goods/01/2077/078


Name of The Project: Supply, Delivery, Installation, Commissioning and Testing of 128 Slice Computed Tomography (CT) Scan Machine with complete set of accessories at Nepal Korea Friendship Municipality Hospital

Date of Bid Publication: 2021/04/20

The Following Technical Specification of Amended Technical Specification of 128 slice Computed Tomography (CT) Scanner have been Amended.

Amended Technical Specification of 128 slice Computed Tomography (CT) Scanner

S.N.	Technical Specification	Bidder's Proposed Specifications	Page no. in technical datasheet	Deviation (if any)
4.3.1	Entire range of rotation times for full 360 degree shall be 0.35 seconds or less for excellent cardiac acquisition & whole body applications.			
4.11.1	It shall have High speed CPU with the latest generation processor with configuration of 16 GB RAM, 1TB hard drive or better configuration. The system shall have all functions menu driven. All applications like scanning, image reconstruction, filming, MPR, CT Angiography, maximum intensity projection MIP, 3D Volume rendering technique, Virtual Endoscopy must be possible from the console. The image reconstruction time must be 30 images/second or more in 512 x 512 matrix or more or better.			


कृष्ण गोपाल श्रेष्ठ
इन्जिनियर